

ADDENDUM No. 02		LSU	BID DUE DATE AND TIME	
BOARD OF SUPERVISORS OF LOUISIANA STATE UNIVERSITY AND AGRICULTURAL & MECHANICAL COLLEGE			05/10/2006	11:00 AM CT
SOLICITATION 000000492 VENDOR # VENDOR NAME AND ADDRESS <div style="border: 1px solid black; height: 80px; width: 350px; margin-top: 10px;"></div>			RETURN ADDENDUM TO FAX # (225)578-2292 LSU LOUISIANA STATE UNIV. PURCHASING OFFICE 213 THOMAS BOYD HALL Baton Rouge LA 70803 BUYER Ashley E. Terrio, CPPB BUYER PHONE (225)578-2097 ISSUE DATE 05/02/2006	

TITLE: STERILIZER

Notice is given to all parties that this Solicitation is amended by the University as stated herein. This Addendum is hereby made an official part of this Solicitation.

Response to Vendor Inquiries: Q) Is installation required of the vendor? A) Yes, installation and removal of existing system is required. Q) Is the unit free-standing with side panels? A) Yes, panels on the sides and top are removable for maintenance. Q) Is the unit to be installed through a wall? A) No, it is free standing Q) Are there existing units present which will require de-installation? A) Yes, an older autoclave of similar construction and weight is in place in Room 1154 EC&E Q) Are there utilities available now and who is responsible for the final hook up? A) Utilities are available.
Successful vendor w

This Addendum is to be signed and returned with your bid or otherwise acknowledged therein. If you have already submitted your bid, and this Addendum creates a need to revise/clarify your original response in any way, you are required to submit such in writing. To be considered, your addendum response must be submitted to and received by the issuing LSU Campus/Department at the "Return Bid To" address stated above no later than the specified bid due date/time. Submittals must be clearly marked with the solicitation number and the bid due date/time and returned via fax, courier service, hand delivery, or USPS mail. Bid revisions received after bid opening cannot be considered, whereupon the bidder must either honor or withdraw its original bid.

Bidder's Addendum Acknowledgement/Response:

As an authorized agent/signatory of the bidder, I/we acknowledge receipt of this Addendum, and
 _____ submit no alterations/clarifications to our original bid.

_____ submit superseding revisions/clarifications to our original bid as written herein or attached hereto.

BIDDER (Name of Firm)	MAILING ADDRESS
AUTHORIZED SIGNATURE	CITY, STATE ZIP
PRINTED NAME	PHONE #
TITLE	FAX #
E-MAIL	FEDERAL TAX ID #

LSU IS AN EQUAL OPPORTUNITY/ACCESS UNIVERSITY

ADDENDUM No. 02	LSU	BID DUE DATE AND TIME		PAGE 2
BOARD OF SUPERVISORS OF LOUISIANA STATE UNIVERSITY AND AGRICULTURAL & MECHANICAL COLLEGE		05/10/2006	11:00 AM	CT
SOLICITATION 000000492				
<p>ill be required to disconnect water and power lines from existing unit and connect to replacement unit. An inspection by LSU Facility Services and possibly the State Fire Marshall will be required because the new autoclave will have an electric steam generator. Q) How much training will be required? A) Minimum one day training on operation.</p>				